



Microdial Flowmeter

Instructions for Use



**UK
CA
0120**

1. Symbols

Warning! Indicates a potentially hazardous situation which, if not avoided, could result in personal injury to the user or others

Caution! Indicates a potentially hazardous situation which, if not avoided, could result in damage to the equipment or property



Use no oil



Service due date

2. Warnings and Cautions

2.1. Warnings!

- ▶ Read through this entire instruction manual before using or showing others how to use a Microdial Flowmeter. As with all medical equipment, attempting to use this device without a thorough understanding of its operation may result in patient or user injury.
- ▶ Medical oxygen is or should be considered a drug and should only be used for medical purposes as prescribed by a physician or authorised clinician and in accordance with the medicinal product labelling.
- ▶ If using a cylinder and regulator, ensure that the device is connected to the regulator and the cylinder valve is properly opened before beginning therapy.
- ▶ Ensure that the medical oxygen supply is sufficient for the proposed therapy and is supplied within the pressure range given in the Device Specification. If the supply is a gas cylinder, check the cylinder contents gauge regularly.
- ▶ Oxygen therapy may be a critical treatment. A regulated flowmeter should be used in strict accordance with the prescription and instructions of a physician. The effectiveness of supplemental oxygen therapy can only be determined by continuous monitoring of blood oxygen levels. It is essential that PaO₂ or SpO₂ monitoring is carried out when supplemental oxygen is prescribed.
- ▶ The Microdial Flowmeter will deliver zero flow between flow settings. When selecting a new flow setting, ensure that the flow-selection dial clicks into place and that a flow is correctly selected.
- ▶ Do not cover the vent hole in the pre-regulator as this will change the device calibration.
- ▶ Do not submerge a Microdial Flowmeter in any fluid. Ensure that no water is allowed to enter ports or vent holes.
- ▶ The accuracy of the flowmeter will be affected if the input pressure is other than stated in the specifications.
- ▶ Always disconnect the flowmeter from the gas supply when it is not in use.
- ▶ Oxygen is not flammable; however an oxygen enriched atmosphere will drastically increase the rate and severity of combustion. Oil and/or grease in the presence of an oxygen enriched atmosphere will become highly combustible. Oxygen must never be allowed to

come into contact with oil, grease or other hydrocarbon based substances. Do not use oil or grease on this Microdial Flowmeter.

- ▶ Many hand creams and moisturisers contain paraffin and petroleum bases which are highly flammable and must never be allowed to contact the Demand Valve. Ensure hands are clean and dry before operating the equipment.
- ▶ Never administer oxygen while smoking or when near an open flame.
- ▶ A gas specific connector is fitted to the Microdial Flowmeter. Do not attempt to modify the fittings to suit other gases or fitting systems.
- ▶ Only appropriately trained personnel working in controlled conditions may disassemble or assemble this Microdial Flowmeter.

2.2. Cautions!

- ▶ The performance of the flowmeter may be affected if it is stored or transported in temperatures outside of the range -20 °C to +60 °C (-4 °F to +140 °F).

3. Functional Description



3.1. Intended Use

The Microdial Flowmeter is intended to control the flow of medical oxygen during oxygen therapy in both homecare and clinical environments. The flow selection dial has 12 positions, including 11 discrete flow rates and an off position. Flow is indicated in litres per minute (l/min) and is visible through a window adjacent to the flow selection dial.

Medical oxygen is or should be considered a drug and should only be used for medical purposes on the authority of a physician and then strictly in accordance with their instructions. Always refer to the medical oxygen product labelling.

3.2. Technical Description

The Microdial Flowmeter has a flow selection dial that is rotated to select the desired oxygen flow. Rotating the dial changes the size of the orifice through which the gas passes and consequently adjusts the gas flow rate.

The Microdial Flowmeter includes a pre-regulator (pressure regulator) upstream of the metering orifice plate. The pre-regulator provides a consistent pressure upstream of the metering orifice plate over a wide range of inlet pressures.

The Microdial Flowmeter has inlet and outlet connectors. The inlet connector is the larger of the two and is a gas specific oxygen probe for connection to the oxygen supply. The smaller outlet connector may be either a barbed outlet for direct connection to oxygen delivery tubing or a threaded DISS (9/16" UNF) connector for connection to a bubble humidifier.

4. Operating Instructions

4.1. Preparation and Connection

Turn the flow selection dial fully clockwise and check that the indication in the flowmeter window is zero.

Connect the medical oxygen probe securely to the oxygen wall outlet or the equivalent outlet of a pressure regulator connected to an oxygen cylinder.

Warning! Where the gas specific connector is dependent on a threaded fastener (e.g. DISS CGA – V5 1240, AS 2902/SIS handwheel) offer device to outlet and connect a few turns. Align device to the final vertical position and fully hand tighten the connection before turning on the supply pressure. Do not use the device to tighten or lock the connection.

For quick connector probes (e.g. BS5682, DIN, AFNOR), ensure that the connection is correctly made by gently pulling the flowmeter before turning on the supply pressure.

If the medical oxygen supply is provided from a gas cylinder, turn on the oxygen supply at the cylinder.

Connect a sufficient length of oxygen tubing (not supplied) to the outlet barb.

4.2. Testing Prior to Use

Turn the flow selection dial to its highest setting and check that gas flow can be felt at the patient end of the supply tubing. If no medical oxygen flow is sensed refer to Section 7 Troubleshooting of this manual.

Turn off the oxygen flow by turning the flow selection dial fully clockwise.

4.3. Operation

Connect the free end of the oxygen tubing to the patient or patient enclosure using the appropriate connector (not supplied).

Determine the approximate oxygen flow rate in l/min required for the patient. Turn the flow selection dial until the closest rate to this is clearly visible in the flowmeter window. Ensure the flow selection dial is in the latched position and the flow rate is visible in the window. It will be obvious to the touch when the selection dial clicks into place.

Warning! The Microdial Flowmeter will deliver zero flow between flow settings.

If the patient requires more or less oxygen flow, this is simply achieved in distinct stages by rotating the flow selection dial.

Check the gas cylinder contents regularly (if applicable) during use of the Microdial Flowmeter and be aware that the supply tubing may be a trip hazard.

4.4. After Use

When therapy is complete, disconnect the Microdial Flowmeter from the gas supply. Where the medical oxygen is being supplied from a cylinder, turn the cylinder off.

Store the Microdial Flowmeter in a clean and dry environment between uses.

5. Maintenance

5.1. Interim Inspection

Microdial Flowmeters should be inspected and tested on an annual basis to ensure correct performance.

Warning! If any defect is found during inspection, the device must be taken out of service.

5.1.1. Inspection

Check the exterior condition of the device and the gas inlet connector. Check that the flow control dial clicks securely into each position. Check the condition of the inlet connector seal if applicable.

5.1.2. Internal Leak Test

Set the flow selection dial to zero and connect to a gas supply. Connect tubing to the outlet barb and immerse the other end of the tube in water. Observe whether any bubbles are formed, which will indicate an internal leak.

5.1.3. Flow Rate Verification

Verify flow rates at all flow settings against those given in the Device Specification in section 6. Mass flow meters with appropriate full-scale ranges for the flow rates are particularly suitable for this.

Alternatively, where flow measuring equipment is not available, a qualitative check can be performed by confirming that gas flow can be sensed to be increasing with each step increase in flow across the full flow range. For higher flow rates, this can be sensed audibly or by holding a hand close to the flow outlet; moistening the skin can increase the cooling sensation and therefore sensitivity to the flow.

For very flow rates, this may better be achieved by placing the end of a connected tube in a glass of water to observe the bubbles.

5.2. Cleaning

Wipe over the external surfaces of the Microdial Flowmeter with an alcohol or disinfecting wipe. Do not allow the ingress of water or other solution into the device.

5.3. Planned Preventative Servicing

The Microdial Flowmeters must be serviced every 2 years to ensure that it continues to perform in accordance with its specification. Microdial Flowmeters have a Service Due date on their labelling adjacent to the spanner / wrench symbol, indicating when the next service is due.

When serviced by BPR Medical Ltd, and where the time to end of life is less than the normal service interval, the Service Due date will be replaced by the End of Life date and preceded by a symbol (⌚). In these cases, the date now indicates when the device reaches end of life.

Warning! Servicing must be carried out by a competent person working in a controlled environment.

Full details of the recommended servicing requirements can be found in the Service Manual. The Service Manual can be obtained from your local BPR Medical distributor, details of which can be found at www.bprmedical.com. Complete the service in accordance with the instructions given in the Service Manual or return the device to a recognised BPR Medical Service Centre on or before the date shown.

6. Specification

Specification	Value
Supply Pressure	Nominal Inlet Pressure 345 kPa to 500 kPa (50 psi to 73 psi)
	Maximum 1000 kPa (145 psi)
	Minimum 280 kPa (40 psi)
Flow Setting (l/min)	Range A: 0, 0.02, 0.03, 0.05, 0.08, 0.12, 0.2, 0.3, 0.5, 0.75, 1.0, 3.0
	Range E: 0, 0.01, 0.02, 0.03, 0.04, 0.05, 0.06, 0.07, 0.08, 0.09, 0.1, 1.0
Filtration	Sintered bronze: 40 µm nominal first stage, 5 µm nominal second stage
Flow Accuracy	±10 % of setting at 1 l/min and above, +20 % to -20 % of setting below 1 l/min
Effects on Accuracy	Varying Inlet Pressure: Less than 6 % of reading for a ±100 kPa change in nominal inlet pressure. Less than 15 % change in reading in the range 280 kPa to 580 kPa.
	Varying Temperature: ±7 % in the range 20 °C ±20 °C
	Varying Outlet Resistance: Less than 1 % of reading up to 5 kPa back pressure
Environmental	Transport/Storage/Operation: -20 °C to +60 °C (-4 °F to +140 °F) Humidity: 0-100% RH non-condensing
Regulatory	UKCA: UK Medical Devices Regulations 2002 – class IIa FDA: Class 1
Applied Standards	
BS EN ISO 10524-4	Pressure regulators for use with medical gases. Low-pressure regulators
BS EN ISO 15001	Anaesthetic and respiratory equipment. Compatibility with oxygen
BS EN ISO 15002	Flow-metering devices for connection to terminal units of medical gas pipeline systems
BS 5682	Specification for probes (quick connectors) for use with medical gas pipeline systems

7. Troubleshooting

Fault	Possible Cause	Solution
No gas flow	Gas cylinder turned off	Check gas supply
	Flowmeter not connected properly	Check gas specific probe is correctly connected
	Gas cylinder empty	Replace gas cylinder
	Medical gas terminal unit on a pipeline system is isolated	Seek advice from someone authorised to operate the medical gas pipeline system isolation valves
	Filter blocked	Service or repair required
	Selection dial set to '0'	Select a positive flow rate
Internal/external leak	Seal failure	Service or repair is required
Insufficient gas flow	Filter partially occluded	Service is required
	Supply pressure too low and/or gas cylinder nearly empty	Check gas supply and/or replace gas cylinder

8. Parts and Spares List

Part Number	Description
816-0001	Microdial Flowmeter - Range A 0.02 - 3 l/min - BS 5682/Barb
816-0003	Microdial Flowmeter - Range E 0.01 - 1 l/min - BS 5682/Barb
816-0006	Microdial Flowmeter - Range A 0.02 - 3 l/min - BS 5682/DISS
816-0011	Microdial Flowmeter - Range E 0.01 - 1 l/min - BS 5682/DISS
816-0026	Microdial Flowmeter - Range A - Hose Barb Outlet (AS 2902)
816-0027	Microdial Flowmeter - Range E - Hose Barb Outlet (AS 2902)
816-0039	Microdial Flowmeter - Range A - Hose Barb Outlet (BS 5682)
816-0040	Microdial Flowmeter - Range E - Hose Barb Outlet (BS 5682)

Spare Parts and Servicing

604-0034	Microdial Flowmeter Output Barb Assembly
604-0035	Microdial Flowmeter DISS Assembly
610-0040	Microdial Flowmeter Service Kit (5 pack)
999-0003	Microdial Flowmeter Service

9. Distributor Details

Australia

BOC Healthcare
Customer Service Centre
10 Julius Avenue
North Ryde
Sydney
NSW, 2113
T: +61 (0)2 8874 4400
F: +61 (0)2 9886 9000
www.boc.com.au

New Zealand

BOC Healthcare
Customer Service Centre
970-992 Great South Road
Penrose 1061
Auckland
New Zealand
T: +64 (0)9 525 5600
www.boc.co.nz

Canada

BOMImed Inc.
96 Terracon Place
Winnipeg
MB R2J 4G&
Canada
T: +1 204.633.2664
F: +1 204.632.1064
www.bomimed.com

BPR Medical Ltd License No: 83602



BPR Medical Limited
22 Hamilton Way, Mansfield, Notts.
NG18 5BU, United Kingdom

+44 (0)1623 628 281
info@bprmedical.com
www.bprmedical.com